

Copper**Code : XL-602****Range : 0.3, 0.6, 1, 1.5, 2 & 3 ppm as Cu.****AQUA-XL**
Water Analysing Kits**Directions for use :**

1. Take 10 ml of sample in the Test Jar.
2. Add 10 drops of **Reagent Cu-1**. Mix well.
3. Add 10 drops of **Reagent Cu-2**. Mix well.
4. Now rinse the empty comparator tube with above solution and fill upto the half level of the comparator tube.
5. Insert this tube in one of the sample slot (S) of the Comparator in such a way that the sample tube colour is close to the matching colour of the comparator.
6. Note down the **ppm level of Copper as Cu. after colour comparison.**